

Connecting via Winsock to STN

Welcome to STN International! Enter x:x

Welcome to STN International! Enter x:
LOGINID:sssptal805jxb

PASSWORD:
TERMINAL (ENTER 1, 2, 3, OR ?):2

* * * * * Welcome to STN International * * * * *

NEWS	1		Web Page URLs for STN Seminar Schedule - N. America
NEWS	2		"Ask CAS" for self-help around the clock
NEWS	3	Jul 12	BEILSTEIN enhanced with new display and select options, resulting in a closer connection to BABS
NEWS	4	AUG 02	IFIPAT/IFIUDB/IFICDB reloaded with new search and display fields
NEWS	5	AUG 02	CAPLUS and CA patent records enhanced with European and Japan Patent Office Classifications
NEWS	6	AUG 02	The Analysis Edition of STN Express with Discover! (Version 7.01 for Windows) now available
NEWS	7	AUG 27	BIOCOMMERCE: Changes and enhancements to content coverage
NEWS	8	AUG 27	BIOTECHABS/BIOTECHDS: Two new display fields added for legal status data from INPADOC
NEWS	9	SEP 01	INPADOC: New family current-awareness alert (SDI) available
NEWS	10	SEP 01	New pricing for the Save Answers for SciFinder Wizard within STN Express with Discover!
NEWS	11	SEP 01	New display format, HITSTR, available in WPIDS/WPINDEX/WPIX
NEWS	12	SEP 14	STN Patent Forum to be held October 13, 2004, in Iselin, NJ
NEWS	13	SEP 27	STANDARDS will no longer be available on STN
NEWS	14	SEP 27	SWETSCAN will no longer be available on STN
NEWS EXPRESS			JULY 30 CURRENT WINDOWS VERSION IS V7.01, CURRENT MACINTOSH VERSION IS V6.0c(ENG) AND V6.0Jc(JP), AND CURRENT DISCOVER FILE IS DATED 11 AUGUST 2004
NEWS HOURS			STN Operating Hours Plus Help Desk Availability
NEWS INTER			General Internet Information
NEWS LOGIN			Welcome Banner and News Items
NEWS PHONE			Direct Dial and Telecommunication Network Access to STN
NEWS WWW			CAS World Wide Web Site (general information)

Enter NEWS followed by the item number or name to see news on that
specific topic.

All use of STN is subject to the provisions of the STN Customer
agreement. Please note that this agreement limits use to scientific
research. Use for software development or design or implementation
of commercial gateways or other similar uses is prohibited and may
result in loss of user privileges and other penalties.

* * * * * STN Columbus * * * * *

FILE 'HOME' ENTERED AT 15:09:53 ON 30 SEP 2004

=> file .pub

COST IN U.S. DOLLARS

FULL ESTIMATED COST

SINCE FILE	TOTAL
ENTRY	SESSION
0.21	0.21

FILE 'MEDLINE' ENTERED AT 15:10:02 ON 30 SEP 2004

FILE 'BIOSIS' ENTERED AT 15:10:02 ON 30 SEP 2004

=> s microtiter and reader

L1 470 MICROTITER AND READER

=> s l1 and software

L2 20 L1 AND SOFTWARE

=> s l2 and py<2000

L3 17 L2 AND PY<2000

=> duplicate remove l3

DUPLICATE PREFERENCE IS 'MEDLINE, BIOSIS'

KEEP DUPLICATES FROM MORE THAN ONE FILE? Y/(N):n

PROCESSING COMPLETED FOR L3

L4 12 DUPLICATE REMOVE L3 (5 DUPLICATES REMOVED)

=> d 1-12 bib ab

L4 ANSWER 1 OF 12 BIOSIS COPYRIGHT (c) 2004 The Thomson Corporation. on
STN

AN 1997:216887 BIOSIS

DN PREV199799523391

TI Rapid evaluation of plant extracts and essential oils for antifungal
activity against Botrytis cinerea.

AU Wilson, C. L. [Reprint author]; Solar, J. M.; Ghaouth, A. El; Wisniewski,
M. E.

CS USDA-ARS Appalachian Fruit Res. Stn., Kearneysville, WV 25430, USA

SO Plant Disease, (1997) Vol. 81, No. 2, pp. 204-210.

CODEN: PLDIDE. ISSN: 0191-2917.

DT Article

LA English

ED Entered STN: 22 May 1997

Last Updated on STN: 22 May 1997

AB A rapid assay to determine antifungal activity in plant extracts and
essential oils is described. Wells in **microtiter** plates were
loaded with Botrytis cinerea spores and plant extracts or essential oils.
Subsequent changes in optical density following spore germination in the
wells was measured after 24 h using an automatic **microtiter**
plate **reader** driven by a **software** program developed
for this purpose. Extracts from 345 plants and 49 essential oils were
evaluated for their antifungal activity against B. cinerea. Among 345
plant extracts analyzed, 13 showed high levels of antifungal activity,
with species of Allium and Capsicum predominating. Among the 49 essential
oils tested, palmarosa (Cymbopogon martini), red thyme (Thymus zygis),
cinnamon leaf (Cinnamomum zeylanicum), and clove buds (Eugenia
caryophyllata) demonstrated the most antifungal activity against B.
cinerea. The most frequently occurring constituents in essential oils
showing high antifungal activity were: D-limonene, cineole; beta-myrcene;
alpha-pinene, beta-pinene; and camphor.

L4 ANSWER 2 OF 12 MEDLINE on STN

AN 1998062498 MEDLINE

DN PubMed ID: 9398357

TI Continuous assay of proteases using a **microtiter** plate
fluorescence **reader**.

AU Menges D A; Ternullo D L; Tan-Wilson A L; Gal S

CS Department of Biological Sciences, State University of New York at
Binghamton, Binghamton, New York 13902-6000, USA.

SO Analytical biochemistry, (1997 Dec 1) 254 (1) 144-7.

Journal code: 0370535. ISSN: 0003-2697.

CY United States

DT Journal; Article; (JOURNAL ARTICLE)

LA English

FS Priority Journals

EM 199801

ED Entered STN: 19980206

Last Updated on STN: 20000303

Entered Medline: 19980127

L4 ANSWER 3 OF 12 MEDLINE on STN DUPLICATE 1
 AN 96089304 MEDLINE
 DN PubMed ID: 8585615
 TI Photometric **microtiter** assay of inorganic phosphate in the presence of acid-labile organic phosphates.
 AU Drueckes P; Schinzel R; Palm D
 CS Theodor-Boveri-Institut für Biowissenschaften (Biocenter) of the University, Würzburg, Germany.
 SO Analytical biochemistry, (1995 Sep 1) 230 (1) 173-7.
 Journal code: 0370535. ISSN: 0003-2697.
 CY United States
 DT Journal; Article; (JOURNAL ARTICLE)
 LA English
 FS Priority Journals
 EM 199603
 ED Entered STN: 19960327
 Last Updated on STN: 19970203
 Entered Medline: 19960321
 AB A simple and rapid colorimetric microassay for inorganic phosphate in the mild acid pH range has enabled us to perform extensive enzyme kinetic studies even in the presence of high concentrations of acid-labile substrates. The assay is performed in a 96-well **microtiter** plate using 30-microliters samples containing between 0.5 and 100 nmol Pi. Compared to existing microassays, color formation is linear over a much wider range of phosphate concentrations and reaction conditions. Using a computer-assisted **microtiter** plate reader, the data can be directly transferred to kinetic analysis **software**. The above assay was used to determine simultaneously up to 12 rates of the Escherichia coli maltodextrin phosphorylase-catalyzed reaction in the direction of glycogen synthesis under conditions which include screening of substrates and inhibitors at variable substrate concentrations and pH on a single 96-well **microtiter** plate.

L4 ANSWER 4 OF 12 MEDLINE on STN
 AN 94035792 MEDLINE
 DN PubMed ID: 7693246
 TI [Partially automated antigen determination and antibody detection with **microtiter** plates].
 Teilautomatisierte Antigenbestimmung und Antikorpersuche auf Mikrotitrationsplatten.
 AU Rapp C; Weisshaar C
 CS Blutspendedienst Hessen des DRK, Institut Kassel, BRD.
 SO Beiträge zur Infusionstherapie = Contributions to infusion therapy, (1993) 31 157-61.
 Journal code: 8812367. ISSN: 1011-6974.
 CY Switzerland
 DT Journal; Article; (JOURNAL ARTICLE)
 LA German
 FS Priority Journals; AIDS
 EM 199312
 ED Entered STN: 19940117
 Last Updated on STN: 19960129
 Entered Medline: 19931203
 AB In addition to several conventional methods for the detection of red cell antigens, the use of microplates has various advantages either as a solid-phase assay (enzyme immunoassay) or as native microplate. Microplates may also be used for the detection of red cell antibodies in 'pooled-cell solid-phase assays' of the second generation and for antibody screening. Blood donors and patients are the two main fields which are to be examined in immunohematology. There are various advantages in using the microplate in blood group serology: (i) if there is hardware already available, like sample processors and microplate **readers**, the use of microplates in blood group serology reduces the costs even if the equipment has to be purchased for this purpose only; (ii) low quantities of reagents are used in microplate assays; (iii) the application of bar codes on tubes and microplates guarantees the most security in sample identification; (iv) it is possible to investigate blood samples selectively depending on the available **software** if antibody

detection is done as the sixth test beside anti-HIV, anti-HCV, HBsAG, lues antibodies and ALT, and (v) recording of data will be easy if electronic data processing is used.

L4 ANSWER 5 OF 12 MEDLINE on STN DUPLICATE 2
AN 91150832 MEDLINE
DN PubMed ID: 2291479
TI **Microtiter** plate assay for the measurement of glutathione and glutathione disulfide in large numbers of biological samples.
AU Baker M A; Cerniglia G J; Zaman A
CS Department of Radiation Oncology, University of Pennsylvania, Philadelphia 19104.
NC CA 4498-03 (NCI)
SO Analytical biochemistry, (1990 Nov 1) 190 (2) 360-5.
Journal code: 0370535. ISSN: 0003-2697.
CY United States
DT Journal; Article; (JOURNAL ARTICLE)
LA English
FS Priority Journals
EM 199104
ED Entered STN: 19910419
Last Updated on STN: 19980206
Entered Medline: 19910402
AB By combining the least complicated and expedient methods of sample handling with the sensitivity and specificity of the GSH assay by enzymatic recycling and the small volumes and **software** capabilities of **microtiter** plate technology we have devised a rapid, sensitive, and easy assay for GSH and GSSG in biological samples. The assay is sensitive to 5 pmol in sample volumes of 50 microliters, although other volumes could be used. The use of a computer-driven microplate with **software** capable of linear kinetic data storage and analysis on each well, Maxline series microplate **readers** and Softmax **software**, enables the user not only to assay large numbers of samples per day but also to have immediate calculated results. We suggest by examples that measurements of total GSH as well as changes in GSH:GSSG in vitro and in vivo are feasible with this technology.

L4 ANSWER 6 OF 12 MEDLINE on STN DUPLICATE 3
AN 90250531 MEDLINE
DN PubMed ID: 2187070
TI Automation and computerization of chromogenic LAL assay method for bacterial endotoxin using 96-well **microtiter** plate.
AU Schadeewald L K; Martin D G; Tsuji K
CS Pharmaceutical Quality Control Division, Upjohn Company, Kalamazoo, Michigan.
SO Journal of parenteral science and technology : a publication of the Parenteral Drug Association, (1990 Mar-Apr) 44 (2) 50-3.
Journal code: 8103145. ISSN: 0279-7976.
CY United States
DT Journal; Article; (JOURNAL ARTICLE)
LA English
FS Priority Journals
EM 199006
ED Entered STN: 19900720
Last Updated on STN: 19900720
Entered Medline: 19900618
AB A high degree of automation was achieved for a chromogenic substrate Limulus ameocyte lysate (LAL) assay method for bacterial endotoxin using a Cetus Pro/Group Liquid Handler and 96-well **microtiter** plates. A Titertek Multiskan **Reader** was interfaced with an IBM PC using a LOTUS MEASURE **software** to capture optical density values of samples in a LOTUS 1-2-3 spreadsheet. A password protected, menu-driven macro programmed in LOTUS 1-2-3 automates the calculation, evaluation of assay parameters, documentation, and generation of a formatted three-page report suitable as a primary record. All assay operations, including testing 19 samples against a four-point standard curve in replicates of four each, the calculation of results, and generation of a report, are completed in less than 40 minutes. The relative standard deviation (RSD) of the assay is approximately 7%, which compares favorably with the robot

automated system.

L4 ANSWER 7 OF 12 MEDLINE on STN
AN 90054316 MEDLINE
DN PubMed ID: 2510548
TI A **microtiter** plate-based system for the semiautomated growth and assay of bacterial cells for beta-galactosidase activity.
AU Menzel R
CS E.I. du Pont de Nemours, Central Research and Development Department, Wilmington, Delaware 19880-0328.
SO Analytical biochemistry, (1989 Aug 15) 181 (1) 40-50.
Journal code: 0370535. ISSN: 0003-2697.
CY United States
DT Journal; Article; (JOURNAL ARTICLE)
LA English
FS Priority Journals
EM 198912
ED Entered STN: 19900328
Last Updated on STN: 19900328
Entered Medline: 19891212
AB The introduction of automated pipetting devices, **microtiter readers**, and microcomputers makes it possible to significantly increase the number of enzyme assays which can be performed as part of the analysis of a biological process. A number of difficulties must be overcome in any such integrated approach based on the **microtiter** plate. Among these are cell lysis, temperature control, the conversion of **microtiter reader** optical density values to standard 1-cm path length values, and data management. The utility of such a scheme can be extended to gene regulation and bacterial genetics studies, if bacterial cell culture techniques can be incorporated into the scheme. This paper addresses these issues in the application of a semiautomated system to the study of the induction of the *gyrA* promoter by treatment (of a *gyrA*-lac operon fusion-containing strain) with a gyrase inhibitor. This system is specific to the requirements of our studies into the modulation of gene expression by DNA relaxation. The general approach, however, can be readily adapted to other studies.

L4 ANSWER 8 OF 12 MEDLINE on STN DUPLICATE 4
AN 89029125 MEDLINE
DN PubMed ID: 3141087
TI Lactobacillus casei microbiological assay of folic acid derivatives in 96-well **microtiter** plates.
AU Horne D W; Patterson D
CS Department of Biochemistry, Vanderbilt University School of Medicine, Nashville, TN 37232.
NC DK32189 (NIDDK)
SO Clinical chemistry, (1988 Nov) 34 (11) 2357-9.
Journal code: 9421549. ISSN: 0009-9147.
CY United States
DT Journal; Article; (JOURNAL ARTICLE)
LA English
FS Priority Journals
EM 198812
ED Entered STN: 19900308
Last Updated on STN: 19970203
Entered Medline: 19881222
AB Microbiological assay is still widely used for estimating folic acid derivatives in serum and other biological samples. We describe here a modification of this procedure involving use of 96-well **microtiter** plates. This procedure, used with modern, computer-interfaced **microtiter-plate readers** and data-reduction **software**, greatly shortens the time and minimizes reagent costs for this assay. Under the conditions of our assay procedures, all folic acid derivatives tested gave equal growth response for Lactobacillus casei. Results for assays for rat liver extracts showed excellent agreement between the standard bioassay and the 96-well procedure.

L4 ANSWER 9 OF 12 MEDLINE on STN DUPLICATE 5
AN 87252300 MEDLINE

DN PubMed ID: 3598197
 TI Rapid data acquisition from a **microtiter** plate fluorescence
reader and applications in kinetic measurements.
 AU Hesford F; Schmitt M; Lazary S
 SO Journal of immunological methods, (1987 Jun 26) 100 (1-2)
 269-79.
 Journal code: 1305440. ISSN: 0022-1759.
 CY Netherlands
 DT Journal; Article; (JOURNAL ARTICLE)
 LA English
 FS Priority Journals
 EM 198708
 ED Entered STN: 19900305
 Last Updated on STN: 19970203
 Entered Medline: 19870807
 AB Programs written in Applesoft BASIC for the rapid acquisition and
 evaluation of data from a commercially available **microtiter**
 plate fluorescence **reader** are presented. Using the data
 acquisition program, the relative fluorescence readings from all 96 wells
 of the **microtiter** plate (one read cycle of the fluorescence
reader) can be stored in each of up to 90 consecutively numbered
 files on a single-sided diskette. A simple timer circuit is described
 which, when used in conjunction with the above program, initiates the
 fluorescence reading process at preset time intervals, thus making
 automatic acquisition of data possible. A further program plots the data
 from consecutive files on the computer monitor and prints a hard copy if
 required. The feasibility of applying the above system and
software to kinetic measurements in enzyme systems is demonstrated
 using methylumbelliferyl phosphate and an alkaline
 phosphatase/immunoglobulin conjugate. In addition, its use in following
 the formation of extracellular hydrogen peroxide by stimulated
 polymorphonuclear leukocytes using horseradish peroxidase-coupled
 oxidation of the fluorescent compound 7-hydroxy-6-methoxy-coumarin
 (scopoletin) is described.

L4 ANSWER 10 OF 12 MEDLINE on STN
 AN 86249587 MEDLINE
 DN PubMed ID: 3013781
 TI A rapid, flexible method for biochemical assays using a **microtiter**
 plate **reader** and a microcomputer. Application for assays of
 protein, Na,K-ATPase and K-p-nitrophenylphosphatase.
 AU Loomis T C; Stahl W L
 SO International journal of bio-medical computing, (1986 May) 18
 (3-4) 183-92.
 Journal code: 0252005. ISSN: 0020-7101.
 CY ENGLAND: United Kingdom
 DT Journal; Article; (JOURNAL ARTICLE)
 LA English
 FS Priority Journals
 EM 198608
 ED Entered STN: 19900321
 Last Updated on STN: 19900321
 Entered Medline: 19860814
 AB A computerized system which greatly accelerates and eases the collection,
 storage, and analysis of data has been applied to several standard
 biochemical assays. The system uses a commercially available
microtiter plate **reader** connected to an Apple IIe
 microcomputer via a standard serial port. Data are transmitted
 automatically from the **reader** to the microcomputer, where they
 can be viewed, printed, further analyzed immediately, or stored on a
 diskette for later retrieval and processing. Some or all of the data may
 be entered manually. The program calculates a linear least squares best
 fit to a standard curve after correcting all data for blanks, then
 determines the quantities of substrate or product contained in each well
 of a **microtiter** plate. Data from two plates may be combined,
 enabling calculation of enzyme specific activities. This system can be
 adapted to any assay whose final step can be performed by a
microtiter plate **reader**. Its use is described for
 determination of protein concentration, Na,K-ATPase activity, and

K-stimulated p-nitrophenylphosphatase activity.

L4 ANSWER 11 OF 12 MEDLINE on STN
AN 85056366 MEDLINE
DN PubMed ID: 6389704
TI A computer program for the evaluation of ELISA data obtained using an automated **microtiter** plate absorbance **reader**.
AU Caulfield M J; Shaffer D
NC AI-18334 (NIAID)
SO Journal of immunological methods, (1984 Nov 30) 74 (2) 205-15.
Journal code: 1305440. ISSN: 0022-1759.
CY Netherlands
DT Journal; Article; (JOURNAL ARTICLE)
LA English
FS Priority Journals
EM 198501
ED Entered STN: 19900320
Last Updated on STN: 19970203
Entered Medline: 19850123
AB A new computer program is described which calculates titers and antibody concentrations from ELISA data. Optical densities are measured in 96-well **microtiter** plates using an automated colorimeter and simultaneously fed into a microcomputer. The data can then be arranged and printed in an 8 X 12 format corresponding to the format of a 96-well **microtiter** plate. The computer program can also compute the titers of samples if the samples are arranged and titrated in one of the suggested formats. In addition, the titers of unknown samples can be automatically compared with the titer of a standard to obtain concentrations. An ELISA designed to measure the concentration of murine antibodies to the cell wall polysaccharide (PnC) extracted from *Streptococcus pneumoniae* was performed to document the use of the program.

L4 ANSWER 12 OF 12 MEDLINE on STN
AN 84089058 MEDLINE
DN PubMed ID: 6418819
TI A computer-based data analysis system for enzyme-linked immunosorbent assays.
AU Slezak T R; Vanderlaan M; Jensen R H
SO Journal of immunological methods, (1983 Dec 16) 65 (1-2) 83-95.
Journal code: 1305440. ISSN: 0022-1759.
CY Netherlands
DT Journal; Article; (JOURNAL ARTICLE)
LA English
FS Priority Journals
EM 198402
ED Entered STN: 19900319
Last Updated on STN: 19900319
Entered Medline: 19840214
AB A computerized system is presented for automating the data collection, processing, and displaying tasks involved in enzyme-linked immunosorbent assays. This system uses a through-the-well absorbance **reader** of **microtiter** plates interfaced to a minicomputer running the UNIX operating system. Optical density in each well of a 96-well **microtiter** plate is recorded as a function of time for up to 10 time points. These data are automatically transmitted to the remote computer. The rate of product formation is then calculated for each well, and a battery of analysis, display, and comparison programs can then be used by the researcher for data presentation. Using the initial rate of reaction as the basis for quantifying enzyme-linked immunosorbent assays focuses on the catalytic property of the enzyme and allows a large dynamic range of the assay on any plate. These programs can be adapted to virtually any mini- or microcomputer with a graphics display or a plotting device. Assuming moderately powerful computing hardware, throughputs of 50 plates per day are easily achieved. The programs work equally well with peroxidase, beta-galactosidase, or alkaline phosphatase conjugated second antibodies, and with whole cell or soluble antigens.